



**बेलौरी नगरपालिका**  
**नगर कार्यपालिकाको कार्यालय**  
**बेलौरी, कंचनपुर**  
**सुदूरपश्चिम प्रदेश, नेपाल**  
**स्थापित २०७३**

पत्र स. : २०८२।८३

च.न. :

मिति : २०८२।१२।२०



**स्वास्थ्य उपकरणको दररेट पेश गर्ने सम्बन्धी सुचना ।**

प्रस्तुत बिषय सम्बन्धमा यस नगरपालिका अन्तर्गत स्वास्थ्य संस्थाहरूमा प्रयोगशाला सेवा तथा सुरक्षित मातृत्व सेवा सञ्चालनका लागि आवश्यक पर्ने स्वास्थ्य उपकरणहरू खरिदका लागि लागत अनुमान तयार गर्ने प्रयोजनका लागि दररेट आवश्यक भएकाले उपकरण सप्लायर्स फर्महरूले डकुमेन्ट सहित तपसिलमा उल्लेखित स्वास्थ्य उपकरणहरूको दररेट ७ दिन भित्र यस नगरपालिकामा पेश गर्न हुन यो सुचना प्रकाशित गरिएको छ ।

तपसिल

s.n.	Name of Equipment	Specification	Unit	Quantity	Rate	Total
1	Tablet ultrasound machine	<ul style="list-style-type: none"><li>• System shall be supplied with main unit, 1 probes.</li><li>• System should have Multiple imaging modes: Includes B-mode, C-mode, M-mode and pulse Doppler, ensuring accurate visualizations for various clinical applications</li><li>• Should have Android compatibility and Type- C cable or wireless connectivity options.</li><li>• scanning depth must be at least 30 cm or more</li><li>• Must submit ISO 13485: 2003/AC: 2007 for Medical Devices</li><li>• Must submit CE and USFDA approved product certificate</li><li>• warranty 2yrs.</li></ul>	Pec	1		
2	Infant radiant warmer	<ul style="list-style-type: none"><li>• It should have microprocessor based servo Controlled</li><li>• It should have a self test function performed at power on and continuously during operation</li><li>• It should have Safety auto cut off for protection if the temperature exceeds the desired value by 2°C &amp; restarts only when temperature falls back into 2°C</li><li>• European CE or USFDA approved product certificate</li><li>• warranty 2yrs.</li></ul>	pec	1		

नि. प्रमुख प्रशासकीय अधिकृत

3	Delivery bed	<ul style="list-style-type: none"> <li>• Delivery Bed with complete attachments and accessories.</li> <li>• It must have adjustments for height and back positions</li> <li>• It must have collapsible side rails</li> <li>• It must have three sectional mattress and seat section must have large perennial cut.</li> <li>• Must have push grip handles.</li> <li>• Must submit ISO13485 or ISO 9001 for Medical Devices</li> <li>• Comprehensive warranty for 2 years.</li> </ul>	pec	1		
4	Steel stool	<ul style="list-style-type: none"> <li>• Shall be supplied complete with All required fittings and accessories for complete operation</li> <li>• The frame shall be made of epoxy powder-coated mild steel or stainless steel with anti-corrosive finish.</li> <li>• The stool shall be made of high-quality laminated board, ABS plastic, or equivalent material with round shape.</li> <li>• (Minimum Acceptable)Height: 450–600 mm Seat Diameter: 250–300 mm</li> <li>• Base Width: Appropriate to ensure stability (Tolerances <math>\pm 10\%</math> acceptable)</li> </ul>	Pec	1		
5	Digital Microscope	<ul style="list-style-type: none"> <li>• Binocular Microscope (LED) with all the necessary adapters and power cords.</li> <li>• System complete with LED illumination system and research quality optics is required.</li> <li>• Fine focus knob: 2.5<math>\mu</math>m.</li> <li>• Coarse movement stroke: 15mm.</li> <li>• Total travel range is 76mm x 30mm.</li> <li>• Stage height movement by wire guide.</li> <li>• Upper limit stopper.</li> <li>• Torque adjustable on coarse focus.</li> <li>• Adjustment knob.</li> <li>• Must submit ISO 9001 or ISO 13485:2003/AC: 2007 AND CE or USFDA approved product certificate.</li> <li>• Comprehensive warranty for 2 years after installation</li> </ul>	pec	1		
6	Automatic Haematology Analyzer(3 part differential)	<ul style="list-style-type: none"> <li>• Fully automated 3 parts differential haematology analyser.</li> <li>• The instrument should provide at least 22 parameters</li> <li>• Should have touch screen at least 7-inch or more</li> <li>• CE marked or US FDA approved</li> </ul>	Pec	1		
7	Semi Auto Biochemestry Analyzer	<ul style="list-style-type: none"> <li>• Display must be TFT.</li> <li>• Qwerty alphanumeric membrane keyboard</li> <li>• system shall use a UV quartz halogen lamp of 6V/10W as the light source.</li> <li>• reaction volume range between 350 <math>\mu</math>L to 5000 <math>\mu</math>L.</li> <li>• Inbuilt Printer</li> <li>• Ce or USFDA approved.</li> <li>• Automatic self-test feature with LED prompts.</li> <li>• CE or USFDA approved</li> </ul>	Pec	1		
8	Auto Colorimeter	<ul style="list-style-type: none"> <li>• microprocessor-based digital auto colorimeter</li> <li>• provide direct digital readout of absorbance with high accuracy</li> <li>• built-in multiple filters (minimum 9 filters) with long operational life</li> <li>• LED display (7-segment or better)</li> <li>• CE marked or US FDA approved</li> <li>• lated arm for precise positioning.</li> <li>• Ce or USFDA approved</li> </ul>	Pec	1		







नि. प्रमुख प्रशासकीय अधिकृत